

ตารางแสดงวงเงินงบประมาณที่ได้รับจัดสรรและรายละเอียดค่าใช้จ่าย
การจัดซื้อจัดจ้างที่มีไขงานก่อสร้าง

๑. ชื่อโครงการ ชื่อชุดตรวจ NMR probe จำนวน ๑ ชุด โดยวิธีเฉพาะเจาะจง.....
๒. หน่วยงานเจ้าของโครงการ สถาบันวิจัยแสงซินโครตรอน (องค์การมหาชน) สังกัดฝ่ายระบบลำเลียงแสง.....
๓. วงเงินงบประมาณที่ได้รับจัดสรร ๑,๕๐๐,๐๐๐.๐๐ บาท
๔. วันที่กำหนดราคากลาง (ราคาอ้างอิง) ณ วันที่ ๑๓ ธันวาคม ๒๕๖๗ เป็นเงิน CHF ๓๖,๒๓๕.๐๐.....
ราคา/หน่วย (ถ้ามี) ตามตารางราคากลาง คิดเป็นเงินไทยเป็นจำนวนเงินประมาณ ๑,๓๙๗,๒๒๑.๖๐ บาท
ตามอัตราแลกเปลี่ยน บมจ.ไทยพาณิชย์ ลงวันที่ ๑๓ ธันวาคม ๒๕๖๗ (CHF ๑ = ๓๘.๕๖ บาท)
๕. แหล่งที่มาของราคากลาง (ราคาอ้างอิง)
ใบเสนอราคา บริษัท Metrolab Technology SA ประเทศสวิตเซอร์แลนด์.....
๖. รายชื่อเจ้าหน้าที่ผู้กำหนดราคากลาง
 - ๖.๑ นายศรัณยู ไชยช่วย.....
 - ๖.๒ นางสาวนฤมล โม้ทอง.....
 - ๖.๓ ดร.ชาตรี ไสยสมบัติ.....

หมายเหตุ :

แหล่งที่มาของราคากลาง (ราคาอ้างอิง) พิจารณาตามพระราชบัญญัติการจัดซื้อจัดจ้างและการบริหารพัสดุภาครัฐ พ.ศ. ๒๕๖๐ มาตรา ๔ “ราคากลาง” หมายความว่า ราคาเพื่อใช้เป็นฐานสำหรับเปรียบเทียบราคาที่ผู้ยื่นข้อเสนอได้ยื่นเสนอไว้ ซึ่งสามารถจัดซื้อจัดจ้างได้จริง

พิจารณาราคากลาง ตามหลักเกณฑ์ข้อ (๔) ราคาที่ได้มาจากการสืบราคาจากท้องตลาด โดยพิจารณาจากใบเสนอราคาตามท้องตลาดซึ่งมีผู้เสนอราคาและใบเสนอราคามาจำนวน ๑ ราย มีคุณสมบัติตรงตามสถาบันฯ กำหนด



Terms of Reference : TOR

NMR probe Quantity 1 set

1. Technical Specifications

Purchase 1 set of NMR probe, consisting of 4 Unit of equipment as shown in the following table.

No.	Product name	Details	Quantity
1	PT2026 NMR Teslameter Bench Mounting with LabView acquisition software	- Frequency range : 1 MHz – 1 GHz - Resolution : +- 0.1 Hz - Accuracy : +- 5 ppm - Measurement rate : Up to 33 Hz	1 Unit
2	1326-10R-0.46-1.50 1326 pulsed-wave probe for 0.46T to 1.50T, Rubber sample	- Range minimum : 0.46T - Range maximum : 1.5T - Sample material : Rubber	1 Unit
3	1326-10R-0.90-3.20 1326 pulsed-wave probe for 0.90T to 3.2T, Rubber Sample	- Range minimum : 0.90T - Range maximum : 3.2T - Sample material : Rubber	1 Unit
4	THM1176-MF-6M-PC-2m Desktop kit with Medium- Field 3-Axis Hall Magnetometer including : with 2 m long thin cable to sensor	- Sensor cable length 2 m - 3-Axis Medium-Field Hall Magnetometer with 6 meter cable - Full-range calibration on 0.1, 0.3, 1 and 3 T ranges - USB Stick with acquisition software for Windows and Mac, API, and user's manual - Software license from Metrolab including free updates	1 Unit

2. Delivery time

Within 120 days

3. Warranty

5 years

Signature..... Purchaser